

Title: **Quality Control Record Sheet**
Insulin Ab 3 QC Data

Prepared by: Geoff Flood
Approved by: Janet Bradbury

Document: wi/qcrs/0.08b
Revision: 2
Date: 09.01.14
Page 1 of 1

Tracer iodination date: 07 September 2023

Kit lot number: 3INK335B

Kit expiry date: 03 November 2023

125I - Insulin	Lot No. IT335R	Expiry Date: 03 NOV 2023
Assay Buffer	Lot No. MIAB146	Expiry Date: 01 JAN 2025
Anti-human IgG	Lot No. 184PG	Expiry Date: 26 DEC 2024
Calibrators	Lot No. IMS37E	Expiry Date: 25 MAY 2024
Negative Control	Lot No. IN131	Expiry Date: 06 JUL 2024
Kit Controls	Lot No. IMCa37E/IMCb37E	Expiry Date: 25 MAY 2024

Total cpm in 25 µl 125I - Insulin 30143

<u>Sample</u>	<u>% 125I - Insulin</u> <u>bound</u>	<u>u/mL</u>	<u>Range</u> <u>u/mL</u>
Negative control	0.9		
INS-Ab calibrator:			
0.4 u/mL	2.1		
1 u/mL	3.5		
10 u/mL	15.9		
50 u/mL	53.0		
Control I	3.9	1.4	0.5 – 1.6
Control II	19.7	12.5	7 - 17

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies at the time of testing.

Assay Date: 12 Sep 2023

Performed by: J DAVIES

Signature: 

Position: Senior Scientist

Authorised by: DR G FLOOD

Signature: 

Position: Head of Quality Control

Date: 12TH SEP 2023