

Title:

Quality Control Record Sheet

AChRAb Standard Curve QC Data

Document:

wi/qcrs/0.14SCb

Revision:

3

Date:

6 Sep 2018

Page

1 of 1

Prepared by: Geoff Flood

Approved by: Rachael Price

Re: AChRAb Standard Curve (lot no. ASC20u; expiry date. 03 Jun 2024)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ^{125}I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA340** are as follows:

Standard	Actual value at QC	
1	0.29	nmol/litre
2	1.1	nmol/litre
3	4.6	nmol/litre
4	8.8	nmol/litre
Value of positive control 116PCaf read off standard curve = 4.7 nmol/litre		
(Range: 2.8-6.4 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 01 Sep 2023Performed by: A.LeeSignature: pp. A.LeePosition: Principal TechnicianAuthorised by: DR G floodSignature: [Signature]Position: Head of Quality ControlDate: 4th SEP 2023