

Title: **Quality Control Record Sheet**
AChRAb Standard Curve QC Data

Prepared by: Geoff Flood
Approved by: Rachael Price

Document: **wi/qcrs/0.14SCb**
Revision: **3**
Date: **6 Sep 2018**
Page **1 of 1**

Re: AChRAb Standard Curve (lot no. ASC20w; expiry date. 29 Jul 2024)

To use this standard curve, plot the mean cpm bound for each standard obtained in your assay against the corresponding nmol/litre value shown in the table below.

No ¹²⁵I decay correction is needed.

Our QC data for the standard curve, obtained with AChR lot **RBA342** are as follows:

Standard	Actual value at QC	
1	0.25	nmol/litre
2	0.91	nmol/litre
3	3.8	nmol/litre
4	9.1	nmol/litre
Value of positive control 116PCah read off standard curve = 4.7 nmol/litre		
(Range: 2.8-6.4 nmol/litre)		

Materials of human origin used in the manufacture of this product have been tested and found non-reactive for HIV1 and 2 and HCV antibodies and HBsAg at the time of testing.

Assay Date: 26 Oct 2023
Performed by: A Lee
Signature: 
Position: Principal Technician

Authorised by: SARAH NUTE
Signature: 
Position: pp Head of Quality Control
Date: 27 OCT 2023